

HUMAN CLARA CELL PROTEIN ELISA

Product Data Sheet

Cat. No.: RD191022200

European Union:  

Rest of the world:
For research use only!

CONTENTS

1.	INTENDED USE	3
2.	STORAGE, EXPIRATION	3
3.	INTRODUCTION	4
4.	TEST PRINCIPLE	4
5.	PRECAUTIONS	5
6.	TECHNICAL HINTS	5
7.	REAGENT SUPPLIED	6
8.	MATERIAL REQUIRED BUT NOT SUPPLIED	6
9.	PREPARATION OF REAGENTS	7
10.	PREPARATION OF SAMPLES	9
11.	ASSAY PROCEDURE	9
12.	CALCULATIONS	10
13.	PERFORMANCE CHARACTERISTICS	10
14.	DEFINITION OF THE STANDARD	10
15.	PRELIMINARY POPULATION AND CLINICAL DATA	10
16.	METHOD COMPARISON	10
17.	TROUBLESHOOTING AND FAQs	10
18.	REFERENCES	10
19.	EXPLANATION OF SYMBOLS	10

**»» This kit is manufactured by:
BioVendor – Laboratorní medicína a.s.**

»» Use only the current version of Product Data Sheet enclosed with the kit!

1. INTENDED USE

The RD191022200 Human Clara Cell Protein ELISA is a sandwich enzyme immunoassay for the quantitative measurement of human Clara cell protein.

»» Features

- **European Union: for in vitro diagnostic use**
Rest of the world: for research use only!
- The total assay time is less than 4 hours
- The kit measures Clara cell protein in serum and plasma (EDTA, citrate, heparin)
- Assay format is 96 wells
- Quality Controls are human serum based
- Standard is recombinant protein (E.coli) based
- Components of the kit are provided ready to use, concentrated or lyophilized

2. STORAGE, EXPIRATION

Store the complete kit at 2-8°C. Under these conditions, the kit is stable until the expiration date (see label on the box).

For stability of opened reagents see Chapter 9.

3. INTRODUCTION

Human Clara cell protein (CC16, CC10, uteroglobin, urinary protein 1 or Clara cell secretory protein) is a member of the secretoglobulin family of proteins and is a secreted product of non-ciliated bronchiolar Clara cells. Its function remains to be fully elucidated but there is convincing data suggesting its role as an immune-modulating and anti-inflammatory agent. Clara cell protein inhibits phospholipase A2 activity as well as interferon gamma signaling and Th1 vs. Th2 lymphocyte regulation.

Clara cell protein concentrations have been determined in serum, plasma and bronchoalveolar lavage fluid in numerous studies since 1994. In serum, its increase is associated with age and asbestos, nitrogen chloride and ozone exposure. Higher levels of CC16 were demonstrated in patients with sarcoidosis, pulmonary fibrosis and high PEEP ventilation. Decreased serum CC16 levels are found after pulmonary resection in smokers and in subjects with chronic obstructive pulmonary disease, asthma or silica exposure.

Decreased CC16 concentrations were also found in the amniotic fluid of fetuses suffering from pulmonary hypoplasia caused by various mechanisms (diaphragmatic hernia, diabetic fetopathy, Turner and Down syndrome). In pleural effusions, the CC16 concentration appears to be associated with its diffusion from the lung as evidenced by high CC16 levels in cardiac pleural congestion.

Based on the above reports Clara cell protein might be perspective useful diagnostic marker of pulmonary diseases and injuries.

Clinical use and areas of investigation:

Pneumonia and bronchopneumonia

Chronic obstructive pulmonary disease, sarcoidosis, pulmonary fibrosis

Acute lung injury

Asthma and allergic rhinitis

Lung cancer

IgA-nephropathy

4. TEST PRINCIPLE

In the BioVendor Human Clara Cell Protein ELISA, standards, quality controls and samples are incubated in microplate wells pre-coated with polyclonal anti-human Clara cell antibody. After 60 minutes incubation and washing, biotin labelled polyclonal anti-human Clara cell protein antibody is added and incubated with captured Clara cell protein for 60 minutes. After another washing, streptavidin-horseradish peroxidase conjugate is added. After 60 minutes incubation and the last washing step, the remaining conjugate is allowed to react with the substrate solution (TMB). The reaction is stopped by addition of acidic solution and absorbance of the

resulting yellow product is measured. The absorbance is proportional to the concentration of Clara cell protein. A standard curve is constructed by plotting absorbance values against concentrations of standards, and concentrations of unknown samples are determined using this standard curve.

5. PRECAUTIONS

- **For professional use only**
- Wear gloves and laboratory coats when handling immunodiagnostic materials
- Do not drink, eat or smoke in the areas where immunodiagnostic materials are being handled
- This kit contains components of human origin. These materials were found non-reactive for HBsAg, HCV antibody and for HIV 1/2 antigen and antibody. However, these materials should be handled as potentially infectious, as no test can guarantee the complete absence of infectious agents
- This kit contains components of animal origin. These materials should be handled as potentially infectious
- Avoid contact with the acidic Stop Solution and Substrate Solution, which contains hydrogen peroxide and tetramethylbenzidine (TMB). Wear gloves and eye and clothing protection when handling these reagents. Stop and/or Substrate Solutions may cause skin/eyes irritation. In case of contact with the Stop Solution and the Substrate Solution wash skin/eyes thoroughly with water and seek medical attention, when necessary
- The materials must not be pipetted by mouth

6. TECHNICAL HINTS

- Reagents with different lot numbers should not be mixed
- Use thoroughly clean glassware
- Use deionized (distilled) water, stored in clean containers
- Avoid any contamination among samples and reagents. For this purpose, disposable tips should be used for each sample and reagent
- Substrate Solution should remain colourless until added to the plate. Keep Substrate Solution protected from light
- Stop Solution should remain colourless until added to the plate. The colour developed in the wells will turn from blue to yellow immediately after the addition of the Stop Solution. Wells that are green in colour indicate that the Stop Solution has not mixed thoroughly with the Substrate Solution
- Dispose of consumable materials and unused contents in accordance with applicable national regulatory requirements

7. REAGENT SUPPLIED

<i>Kit Components</i>	<i>State</i>	<i>Quantity</i>
Antibody Coated Microtiter Strips	ready to use	96 wells
Biotin Labelled Antibody	ready to use	13 ml
Streptavidin-HRP Conjugate	ready to use	13 ml
Master Standard	lyophilized	2 vials
Quality Control HIGH	lyophilized	2 vials
Quality Control LOW	lyophilized	2 vials
Dilution Buffer	ready to use	20 ml
Wash Solution Conc. (10x)	concentrated	100 ml
Substrate Solution	ready to use	13 ml
Stop Solution	ready to use	13 ml
Product Data Sheet + Certificate of Analysis	-	1 pc

8. MATERIAL REQUIRED BUT NOT SUPPLIED

- Deionized (distilled) water
- Test tubes for diluting samples
- Glassware (graduated cylinder and bottle) for Wash Solution (Dilution Buffer)
- Precision pipettes to deliver 5-1000 μ l with disposable tips
- Multichannel pipette to deliver 100 μ l with disposable tips
- Absorbent material (e.g. paper towels) for blotting the microtitre plate after washing
- Vortex mixer
- Orbital microplate shaker capable of approximately 300 rpm
- Microplate washer (optional). [Manual washing is possible but not preferable.]
- Microplate reader with 450 ± 10 nm filter, preferably with reference wavelength 630 nm (alternatively another one from the interval 550-650 nm)
- Software package facilitating data generation and analysis (optional)

9. PREPARATION OF REAGENTS

- All reagents need to be brought to room temperature prior to use
 - Always prepare only the appropriate quantity of reagents for your test
 - Do not use components after the expiration date marked on their label
- Assay reagents supplied ready to use:

Antibody Coated Microtiter Strips

Stability and storage:

Return the unused strips to the provided aluminium zip-sealed bag with desiccant and seal carefully. Remaining Microtiter Strips are stable 3 months when stored at 2-8°C and protected from the moisture.

Biotin Labelled Antibody

Streptavidin-HRP Conjugate

Dilution Buffer

Substrate Solution

Stop Solution

Stability and storage:

Opened reagents are stable 3 months when stored at 2-8°C.

- Assay reagents supplied concentrated or lyophilized:

Human Clara cell protein Master Standard

Refer to the Certificate of Analysis for current volume of Dilution Buffer needed for reconstitution of standard!!!

Reconstitute the lyophilized Master Standard with Dilution Buffer just prior to the assay. Let it dissolve at least 15 minutes with occasional gentle shaking (not to foam). The resulting concentration of the human Clara cell protein in the stock solution is **50 ng/ml**.

Prepare set of standards using Dilution Buffer as follows:

<i>Volume of Standard</i>	<i>Dilution Buffer</i>	<i>Concentration</i>
Stock	-	50 ng/ml
150 µl of stock	150 µl	25 ng/ml
150 µl of 25 ng/ml	150 µl	12.5 ng/ml
150 µl of 12.5 ng/ml	150 µl	6.25 ng/ml
150 µl of 6.25 ng/ml	150 µl	3.13 ng/ml
150 µl of 3.13 ng/ml	150 µl	1.57 ng/ml

Dilute prepared standards (50 – 1.57 ng/ml) 25x with Dilution Buffer just prior to the assay, e.g. 10 µl of Standard + 240 µl of Dilution Buffer for duplicates. **Mix well** (not to foam). Vortex is recommended.

Stability and storage:

Standard stock solution (50 ng/ml) should be aliquoted and frozen at –20°C for 3 months. Avoid repeated freeze/thaw cycles.

Do not store the diluted Standard solutions.

Quality Controls HIGH, LOW

Refer to the Certificate of Analysis for current volume of distilled water needed for reconstitution and for current Quality Control concentration!!!

Reconstitute each Quality Control (HIGH and LOW) with distilled water just prior to the assay. Let it dissolve at least 15 minutes with occasional gentle shaking (not to foam).

Dilute reconstituted Quality Controls 25x with Dilution Buffer, e.g. 5 µl of Quality Control + 120 µl of Dilution Buffer when assaying samples in singlets, or preferably 10 µl of Quality Control + 240 µl of Dilution Buffer for duplicates. **Mix well** (not to foam). Vortex is recommended.

Stability and storage:

The reconstituted Quality Controls must be used immediately or aliquoted and frozen at -20°C for 3 months. Avoid repeated freeze/thaw cycles.

Do not store the diluted Quality Controls.

Note:

Concentration of analyte in Quality Controls need not be anyhow associated with normal and/or pathological concentrations in serum or another body fluid. Quality Controls serve just for control that the kit works in accordance with PDS and CoA and that ELISA test was carried out properly.

Wash Solution Conc. (10x)

Dilute Wash Solution Concentrate (10x) ten-fold in distilled water to prepare a 1x working solution. Example: 100 ml of Wash Solution Concentrate (10x) + 900 ml of distilled water for use of all 96-wells.

Stability and storage:

The diluted Wash Solution is stable 1 month when stored at 2-8°C. Opened Wash Solution Concentrate (10x) is stable 3 months when stored at 2-8°C.

10. PREPARATION OF SAMPLES

The kit measures Clara cell protein in serum and plasma (EDTA, citrate, heparin).

Samples should be assayed immediately after collection or should be stored at -20°C. Mix thoroughly thawed samples just prior to the assay and avoid repeated freeze/thaw cycles, which may cause erroneous results. Avoid using hemolyzed or lipemic samples.

Dilute serum or plasma samples 25x with Dilution Buffer just prior to the assay, e.g. 5 µl of sample + 120 µl of Dilution Buffer for singlets, or preferably 10 µl of sample + 240 µl of Dilution Buffer for duplicates. **Mix well** (not to foam). Vortex is recommended.

Stability and storage:

Samples should be stored at -20°, or preferably at -70°C for long-term storage. Avoid repeated freeze/ thaw cycles.

Do not store the diluted samples.

See Chapter 13 for stability of serum and plasma samples when stored at 2-8°C, effect of freezing/thawing and effect of sample matrix (serum/plasma) on the concentration of Clara cell protein.

Ask for information at info@biovendor.com if assaying bronchoalveolar lavage fluid or urine.

Note: It is recommended to use a precision pipette and a careful technique to perform the dilution in order to get precise results!

11. ASSAY PROCEDURE

1. Pipet **100 µl** of diluted Standards, Quality Controls, Dilution Buffer (=Blank) and samples, preferably in duplicates, into the appropriate wells. See *Figure 1* for example of work sheet.
2. Incubate the plate at room temperature (ca. 25°C) for **1 hour**, shaking at ca. 300 rpm on an orbital microplate shaker.
3. Wash the wells 3-times with Wash Solution (0.35 ml per well). After final wash, invert and tap the plate strongly against paper towel.
4. Add **100 µl** of Biotin Labelled Antibody into each well.
5. Incubate the plate at room temperature (ca. 25°C) for **1 hour**, shaking at ca. 300 rpm on an orbital microplate shaker.

6. Wash the wells 3-times with Wash Solution (0.35 ml per well). After final wash, invert and tap the plate strongly against paper towel.
7. Add **100 µl** of Streptavidin-HRP Conjugate into each well.
8. Incubate the plate at room temperature (ca. 25°C) for **1 hour**, shaking at ca. 300 rpm on an orbital microplate shaker.
9. Wash the wells 3-times with Wash Solution (0.35 ml per well). After final wash, invert and tap the plate strongly against paper towel.
10. Add **100 µl** of Substrate Solution into each well. Avoid exposing the microtiter plate to direct sunlight. Covering the plate with e.g. aluminium foil is recommended.
11. Incubate the plate for **10 minutes** at room temperature. The incubation time may be extended [up to 20 minutes] if the reaction temperature is below than 20°C. Do not shake the plate during the incubation.
12. Stop the colour development by adding **100 µl** of Stop Solution.
13. Determine the absorbance of each well using a microplate reader set to 450 nm, preferably with the reference wavelength set to 630 nm (acceptable range: 550 – 650 nm). Subtract readings at 630 nm (550 - 650 nm) from the readings at 450 nm.
The absorbance should be read within 5 minutes following step 12.

Note: If some samples and standard/s have absorbances above the upper limit of your microplate reader, perform a second reading at 405 nm. A new standard curve, constructed using the values measured at 405 nm, is used to determine Clara cell protein concentration of off-scale standards and samples. The readings at 405 nm should not replace the readings for samples that were “in range” at 450 nm.

Note 2: Manual washing: Aspirate wells and pipet 0.35 ml Wash Solution into each well. Aspirate wells and repeat twice. After final wash, invert and tap the plate strongly against paper towel. Make certain that Wash Solution has been removed entirely.

	strip 1+2	strip 3+4	strip 5+6	strip 7+8	strip 9+10	strip 11+12
A	Standard 50	QC LOW	Sample 8	Sample 16	Sample 24	Sample 32
B	Standard 25	Sample 1	Sample 9	Sample 17	Sample 25	Sample 33
C	Standard 12.5	Sample 2	Sample 10	Sample 18	Sample 26	Sample 34
D	Standard 6.25	Sample 3	Sample 11	Sample 19	Sample 27	Sample 35
E	Standard 3.13	Sample 4	Sample 12	Sample 20	Sample 28	Sample 36
F	Standard 1.57	Sample 5	Sample 13	Sample 21	Sample 29	Sample 37
G	Blank	Sample 6	Sample 14	Sample 22	Sample 30	Sample 38
H	QC HIGH	Sample 7	Sample 15	Sample 23	Sample 31	Sample 39

Figure 1: Example of a work sheet.

12. CALCULATIONS

Most microplate readers perform automatic calculations of analyte concentration. The standard curve is constructed by plotting the mean absorbance (Y) of Standards against the known concentration (X) of Standards in logarithmic scale, using the four-parameter algorithm. Results are reported as concentration of Clara cell protein ng/ml in samples.

Alternatively, the *logit log* function can be used to linearize the standard curve, i.e. *logit* of the mean absorbance (Y) is plotted against *log* of the known concentration (X) of Standards.

Samples, Quality Controls and Standards are all diluted 25x prior to analysis, so there is no need to take this dilution factor into account.




Figure 2: Typical Standard Curve for Human Clara Cell Protein ELISA.

13. PERFORMANCE CHARACTERISTICS

»» Typical analytical data of BioVendor Human Clara Cell Protein ELISA are presented in this chapter

- **Sensitivity**

Limit of Detection (LOD), defined as concentration of analyte giving absorbance higher than mean absorbance of blank* plus three standard deviations of the absorbance of blank: $A_{\text{blank}} + 3 \times \text{SD}_{\text{blank}}$, is calculated from the real Clara cell protein values in wells and is 46 pg/ml.

*Dilution Buffer is pipetted into blank wells.

- **Limit of assay**

Samples with absorbances exceeding the absorbance of the highest standard should be measured again with higher dilution. The final concentration of samples calculated from the standard curve must be multiplied by the respective dilution factor.

- **Specificity**

The antibodies used in this ELISA are specific for human Clara cell protein with no detectable crossreactivities to the cytokines that may be present in human serum.

Determination of Clara Cell protein does not interfere with hemoglobin (1.0 mg/ml), bilirubin (170 $\mu\text{mol/l}$) and triglycerides (5.0 mmol/l).

Sera of several mammalian species were measured in the assay. See results below.

For details please contact us at info@biovendor.com.

<i>Mammalian serum sample</i>	<i>Observed crossreactivity</i>
Bovine	no
Cat	no
Dog	no
Goat	no
Hamster	no
Horse	no
Monkey	yes
Mouse	yes
Pig	no
Rabbit	no
Rat	no
Sheep	no

- **Precision**

Intra-assay (Within-Run) (n=8)

<i>Sample</i>	<i>Mean (ng/ml)</i>	<i>SD (ng/ml)</i>	<i>CV (%)</i>
1	14.28	0.55	3.82
2	5.31	0.16	2.96

Inter-assay (Run-to-Run) (n=4)

<i>Sample</i>	<i>Mean (ng/ml)</i>	<i>SD (ng/ml)</i>	<i>CV (%)</i>
1	5.37	0.22	4.09
2	7.75	0.49	6.36
3	16.59	0.63	3.78

- **Spiking Recovery**

Serum samples were spiked with different amounts of human Clara cell protein, diluted with Dilution Buffer 25x and assayed.

<i>Sample</i>	<i>Observed (ng/ml)</i>	<i>Expected (ng/ml)</i>	<i>Recovery O/E (%)</i>
1	4.57	-	-
	7.36	7.7	95.6
	10.69	10.8	98.8
	17.37	17.1	101.8
2	6.36	-	-
	8.69	9.5	91.6
	12.28	12.6	97.4
	17.17	18.9	91.0

- **Linearity**

Serum samples (diluted 25x with Dilution Buffer) were serially diluted with Dilution Buffer and assayed.

<i>Sample</i>	<i>Dilution</i>	<i>Observed (ng/ml)</i>	<i>Expected (ng/ml)</i>	<i>Recovery O/E (%)</i>
1	-	10.47	-	-
	2x	5.33	5.2	101.8
	4x	2.87	2.6	109.6
	8x	1.37	1.3	104.8
2	-	13.94	-	-
	2x	6.82	7.0	97.8
	4x	3.76	3.5	107.8
	8x	1.74	1.7	99.9

- **Effect of sample matrix**

EDTA, citrate and heparin plasmas were compared to respective serum samples from the same 10 individuals. Results are shown below:

Volunteer No.	Serum (ng/ml)	Plasma (ng/ml)		
		EDTA	Citrate	Heparin
1	4.74	3.28	4.51	4.71
2	8.61	8.57	6.72	8.63
3	1.94	1.88	1.54	2.02
4	9.06	8.28	6.88	8.56
5	8.24	9.96	7.58	8.67
6	4.77	4.80	4.83	4.53
7	5.87	5.56	4.86	5.93
8	3.95	3.81	3.08	3.76
9	4.08	3.75	3.19	4.06
10	6.49	6.07	5.18	5.91
Mean (ng/ml)	5.77	5.60	4.84	5.68
Mean Plasma/Serum (%)	-	96.9	83.8	98.3
Coefficient of determination R²		0.91	0.93	0.98




Figure 3: Human Clara cell protein levels measured using Human Clara Cell Protein ELISA from 10 individuals using serum, EDTA, citrate and heparin plasma, respectively.

- **Stability of samples stored at 2-8°C**

Samples should be stored at –20°C. However, no decline in concentration of Clara cell protein was observed in serum and plasma samples after 7 days when stored at 2-8°C. To avoid microbial contamination, samples were treated with ε-aminocaproic acid and sodium azide, resulting in the final concentration of 0.03% and 0.1%, respectively.

- **Effect of Freezing/Thawing**

No decline was observed in concentration of human Clara cell protein in serum and plasma samples after repeated (3x) freeze/thaw cycles. However it is recommended to avoid unnecessary repeated freezing/thawing of the samples.

14. DEFINITION OF THE STANDARD

The recombinant protein produced in E. coli is used as the Master Standard in this assay. The Clara cell protein is a 9.2 kDa protein consisting of 80 amino acids.

The CC16 concentration strongly depends on the method, which is used for the protein determination. Master Standard contains 50 ng of CC16 measured by Bradford method (used in this kit), 215 ng of CC16 measured by BCA method and 340 ng of CC16 measured by Lowry method.

15. PRELIMINARY POPULATION AND CLINICAL DATA

The following results were obtained when serum samples from 55 unselected donors (34 men + 21 women) 22 - 61 years old were assayed with the Biovendor Human Clara Cell Protein ELISA in our laboratory:

- **Age and Sex dependent distribution of human Clara Cell Protein**

Sex	Age years	n	Mean	SD	Min	Max	Median
			ng/ml				
Men	20-39	18	6.4	1.9	3.7	9.4	6.2
	40-69	16	7.1	2.8	4.2	14.9	6.4
Women	20-39	12	7.3	2.3	3.6	11.1	7.4
	40-69	9	9.5	4.1	4.5	17.1	8.5




Figure 4: Human Clara Cell Protein concentration plotted against donor age and sex.

- **Typical distribution of Clara cell protein in various body fluids**

Sample	Mean (ng/ml)	Range (ng/ml)
Serum	12.6	3.7 – 23.2
Urine	18.7	0.2 – 88.6
Seminal fluid	1 030.0	145 – 8 600
BAL	1 360.0	154 – 4 300
Synovial fluid	9.1	2.8 – 16.4
Pleural fluid	11.4	0.7 – 32.8
Cerebrospinal fluid	0.5	0 – 5.7
Gastric juice	185.0	0 – 1 220
Bile	0.7	0 – 2.3

Concentrations of Clara cell protein are expressed as ng/ml. See for details:

Shijubo N., Kawabata I., Sato N., Itoh Y.: Clinical Aspects of Clara Cell 10-kDa Protein/ Uteroglobin (Secretoglobin 1A1), Current Pharmaceutical Design, 9, 1139-1149, (2003)

- **Reference range**

It is recommended that each laboratory include its own panel of control sample in the assay. Each laboratory should establish its own normal and pathological reference ranges for Clara cell protein levels with the assay.

16. METHOD COMPARISON

The BioVendor's Human Clara Cell Protein ELISA was compared to the previous version of the ELISA (Cat. No. RD191022200, lot. E12-022). The following correlation graph was obtained.




Figure 5: Method comparison

17. TROUBLESHOOTING AND FAQs

»» Weak signal in all wells

Possible explanations:

- Omission of a reagent or a step
- Improper preparation or storage of a reagent
- Assay performed before reagents were allowed to come to room temperature
- Improper wavelength when reading absorbance

»» High signal and background in all wells

Possible explanations:

- Improper or inadequate washing
- Overdeveloping; incubation time with Substrate Solution should be decreased before addition of Stop Solution

- Incubation temperature over 30°C

»» High coefficient of variation (CV)

Possible explanation:

- Improper or inadequate washing
- Improper mixing Standards, Quality Controls or samples

18. REFERENCES

»» References to Clara cell protein:

- Deraz T, Kamel TB, El-Mogy MI, Moustafa EH. Serum and nasal lavage fluid Clara cell protein decreases in children with allergic rhinitis. *Int J Pediatr Otorhinolaryngol*, 76(9):1241-4 (2012)
- Wuetzler S, Backhaus L, Henrich D, Geiger E: Clara cell protein16: A biomarker for detecting secondary respiratory complications in patients with multiple injuries. *J Trauma Acute Care Surgery* 73(4): 838-842 (2012)
- Chowdhury B, Zhang Z, Mukherjee AB. Uteroglobin interacts with the heparin-binding site of fibronectin and prevents fibronectin-IgA complex formation found in IgA-nephropathy. *FEBS Lett.* 82(5):611-5 (2008)
- Braido F, Riccio AM, Guerra L, Gamalero C, Zolezzi A, Tarantini F, De Giovanni B, Folli C, Descalzi D, Canonica GW. Clara cell 16 protein in COPD sputum: a marker of small airways damage? *Respir Med.* 101(10):2119-24 (2007)
- Shijubo N., Kawabata I., Sato N., Itoh Y.: Clinical Aspects of Clara Cell 10-kDa Protein/ Uteroglobin (Secretoglobin 1A1), *Current Pharmaceutical Design*, 9, 1139-1149, (2003)
- Nord M., Schubert K., Cassel T., Andersson O., Riise G.: Decreased serum and bronchoalveolar lavage levels of Clara cell secretory protein (CC16) is associated with bronchiolitis obliterans syndrome and airway neutrophilia in lung transplant recipients. *Transplantation*, 73, 1264-1269, (2002)
- Petrek M., Hermans C., Kolek V., Fialova J., Bernard A.: Clara cell protein (CC16) in serum and bronchoalveolar lavage fluid of subjects exposed to asbestos. *Biomarkers*, 7(1), 58-67, (2002)
- Hermans C., Petrek M., Kolek V., Weynand B., Pieters T., Lambert M., Bernard A.: Serum Clara cell protein (CC16), a marker of the integrity of the air-blood barrier in sarcoidosis. *Eur Respir J*, 18(3), 507-514 (2001)
- Bernard A., Roels H., Lauwerys R., Witters R., Gielens C., Soumilion A. et al.: Human urinary protein 1: Evidence for identity with the Clara cell protein and occurrence in respiratory tract and urogenital secretions. *Clin Chim Acta*, 207, 239-249, (1992)
- Bernard A., Lauwerys R., Noel A., Vandeleene B., Lambert A.: Urine protein 1: a sex-dependent marker of tubular or glomerular dysfunction. *Clin Chem*, 35, 2141-2142, (198







»» References to this product:

- Taketoshi N, Kazuhito A, Atsuko F, Ken-ichi K: Enhancement of Clara cell 10-kDa protein production from nasal epithelial cells by fexofenadine hydrochloride. *As Pac J Allergy Immunol* 30, 139-145, 2012.
- Bourdin A, Kotsimbos T, Nguyen K, Vachier I, Mainprice B, Farce M, Paganin F, Marty-Ane C, Vernhet H, Godard P, Chanez P. Non-invasive assessment of small airway remodelling in smokers. *COPD*; 7 (2):102-10 (2010)
- Kropski JA, Fremont RD, Calfee CS, Ware LB. Clara cell protein (CC16), a marker of lung epithelial injury, is decreased in plasma and pulmonary edema fluid from patients with acute lung injury. *Chest*; 135 (6):1440-7 (2009)
- Chimenti L, Morici G, Paterno A, Bonanno A, Vultaggio M, Bellia V, Bonsignore MR. Environmental conditions, air pollutants, and airway cells in runners: a longitudinal field study. *J Sports Sci*; 27 (9):925-35 (2009)
- Gaber F, Daham K, Higashi A, Higashi N, Gulich A, Delin I, James A, Skedinger M, Gyllfors P, Nord M, Dahlen SE, Kumlin M, Dahlen B. Increased levels of cysteinyl-leukotrienes in saliva, induced sputum, urine and blood from patients with aspirin-intolerant asthma. *Thorax*; 63 (12):1076-82 (2008)
- Sims MW, Tal-Singer RM, Kierstein S, Musani AI, Beers MF, Panettieri RA, Haczku A. Chronic obstructive pulmonary disease and inhaled steroids alter surfactant protein D (SP-D) levels: a cross-sectional study. *Respir Res*; 9:13 (2008)
- Ulvestad B, Randem BG, Andersson L, Ellingsen DG, Barregard L. Clara cell protein as a biomarker for lung epithelial injury in asphalt workers. *J Occup Environ Med*; 49 (10):1073-8 (2007)
- Coppens JT, Van Winkle LS, Pinkerton KE, Plopper CG . Distribution of Clara Cell Secretory Protein Expression in the Tracheobronchial Airways of Rhesus Monkeys. *Am J Physiol Lung Cell Mol Physiol* . 292:1155-1162 (2007)
- Benson M, Fransson M, Martinsson T, Nalwai AT, Uddman R, Cardell LO . Inverse relation between nasal fluid Clara Cell Protein 16 levels and symptoms and signs of rhinitis in allergen-challenged patients with intermittent allergic rhinitis. *Allergy*. 62(2):178-83 (2007)
- Harvey BG, Heguy A, Leopold PL, Carolan BJ, Ferris B, Crystal RG . Modification of gene expression of the small airway epithelium in response to cigarette smoking. *J Mol Med*. 85(1):39-53 (2007)
- Andersson L, Lundberg PA, Barregard L . Methodological aspects on measurement of Clara cell protein in urine as a biomarker for airway toxicity, compared with serum levels. *J Appl Toxicol*. 27(1):60-66 (2006)
- Schnapp LM, Donohoe S, Chen J, Sunde DA, Kelly PM, Ruzinski J, Martin T, Goodlett DR . Mining the acute respiratory distress syndrome proteome: identification of the insulin-like growth factor (IGF)/IGF-binding protein-3 pathway in acute lung injury. *Am J Pathol*. 169(1):86-95 (2006)
- Fransson M, Adner M, Uddman R, Cardell LO. Lipopolysaccharide-induced down-regulation of uteroglobin in the human nose. *Acta Otolaryngol*. 127(3):285-91 (2006)


- Martin AC, Laing IA, Khoo SK, Zhang G, Rueter K, Teoh L, Taheri S, Hayden CM, Geelhoed GC, Goldblatt J, LeSouef PN . Acute asthma in children: Relationships among CD14 and CC16 genotypes, plasma levels, and severity. *Am J Respir Crit Care Med*. 173(6):617-22 (2006)
- Mattsson J, Remberger M, Andersson O, Sundberg B, Nord M . Decreased serum levels of Clara cell secretory protein (CC16) are associated with bronchiolitis obliterans and may permit early diagnosis in patients after allogeneic stem-cell transplantation. *Transplantation*. 27;79(10):1411-6 (2005)
- Benson M, Jansson L, Adner M, Luts A, Uddman R, Cardell LO . Gene profiling reveals decreased expression of uteroglobin and other anti-inflammatory genes in nasal fluid cells from patients with intermittent allergic rhinitis. *Clin Exp Allergy*. 35(4):473-8 (2005)

»» For more references on this product see our WebPages at www.biovendor.com

19. EXPLANATION OF SYMBOLS

REF	Catalogue number
Cont.	Content
LOT	Lot number
	See instructions for use
	Biological hazard
	Expiry date
	Storage conditions
	Identification of packaging materials
IVD 	In vitro diagnostic medical device


Assay Procedure Summary



12							
11							
10							
9							
8							
7							

Gentaur Molecular Products
Voortstraat 49
1910 Kampenhout, Belgium
<http://www.gentaur-worldwide.com>

5							
4							
3							
2							
1							
	A	B	C	D	E	F	G
							H



Gentaur Molecular Products
Voortstraat 49
1910 Kampenhout, Belgium
<http://www.gentaur-worldwide.com>